



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/084,283

02/26/2002

David R. Diduch

DID-101

7097

39013 7590 04/24/2008  
MOAZZAM & ASSOCIATES, LLC  
7601 LEWINSVILLE ROAD  
SUITE 304  
MCLEAN, VA 22102

EXAMINER

OU, JING RUI

ART UNIT

PAPER NUMBER

3773

MAIL DATE

DELIVERY MODE

04/24/2008

PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

<b>Office Action Summary</b>	<b>Application No.</b> 10/084,283	<b>Applicant(s)</b> DIDUCH ET AL.	
	<b>Examiner</b> JING OU	<b>Art Unit</b> 3773	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 10 January 2008.
- 2a) ☐ This action is **FINAL**.                      2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 50,53,56,57,59-62 and 65-67 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 50,53,56,57,59-62 and 65-67 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☒ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All    b) ☐ Some \*    c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
  - ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- |                                                                                      |                                                                   |
|--------------------------------------------------------------------------------------|-------------------------------------------------------------------|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892)                     | 4) <input type="checkbox"/> Interview Summary (PTO-413)           |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____                                      |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)          | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____                                                          | 6) <input type="checkbox"/> Other: _____                          |

### **DETAILED ACTION**

1. Prosecution on the merits of this application is re-opened, all previous indicated allowable claims are withdrawn. Claims 50, 53, 56, 57, 59-62, and 65- 67 are pending.

#### ***Specification***

2. The disclosure is objected to because of the following informalities: Examiner cannot find support for recitation "the opening in the jaw is partially open" of claims 56, 60, and 65 in the disclosure.

Appropriate correction is required.

#### ***Claim Rejections - 35 USC § 112***

3. The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

4. Claims 56, 60, and 65 are rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention. The recitations, "the opening in the jaw is partially open" in claims 56, 60, and 65 are unclear. Examiner cannot find support for such recitation in the disclosure. In addition, it is unclear and confusing how the jaw is removable from around a side of a strand of suture that has been passed into the soft tissue.

#### ***Claim Rejections - 35 USC § 102***

5. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

6. Claims 50, 59, 60, 62, 65, and 67 are rejected under 35 U.S.C. 102(b) as being anticipated by Caspari et al (US Pat. No.: 4,923,461).

In regard to Claims 50, 59, 60, 62, 65, and 67, Caspari et al discloses a surgical device (suturing instrument, 20, Fig. 1) for passing suture through soft tissue, comprising: a body (elongate tube, 36, Fig. 1) having a lumen to accommodate an elongate member (combination of tubing 104 and needle 56, Figs 5 and 6 and Col. 4, lines 19-22) having a puncturing projection (needle, 56, Fig. 5); a grasping mechanism comprising a first jaw (44, Fig. 2) and second jaw (34, Fig. 2) connected to said body and encompassing at least a portion of the elongate member, one of the jaws moveable relative to the other and having an opening (65, Fig. 1) at the middle region of the jaw to receive at least one strand of suture for allowing the elongate member therethrough and to the soft tissue, and wherein said grasping mechanism clamps soft tissue while the suture passing member slides within the lumen and passes suture through the soft tissue outside of the grasping mechanism through an opening in one of said jaws of said grasping mechanism (Figs. 11-12 and Col. 4, lines 59-68); the puncturing projection (needle, 56, Fig. 5) having a distal portion and carrying a suture (Col. 5, lines 1-5, the needle has an opening at its distal end and carries a suture when the suture passes through the distal opening of the needle), the puncturing projection movable between a first position wherein the distal portion of the puncturing projection is substantially contained within an area between the first jaw and the second jaw and a second position wherein the distal portion of the puncturing projection extends beyond

the first jaw (Col. 3, lines 57 and Figs 1 and 11-15); and a handle (26, Fig. 2) to manipulate movement of the jaws.

***Claim Rejections - 35 USC § 103***

7. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

8. The factual inquiries set forth in *Graham v. John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:

1. Determining the scope and contents of the prior art.
2. Ascertaining the differences between the prior art and the claims at issue.
3. Resolving the level of ordinary skill in the pertinent art.
4. Considering objective evidence present in the application indicating obviousness or nonobviousness.

9. This application currently names joint inventors. In considering patentability of the claims under 35 U.S.C. 103(a), the examiner presumes that the subject matter of the various claims was commonly owned at the time any inventions covered therein were made absent any evidence to the contrary. Applicant is advised of the obligation under 37 CFR 1.56 to point out the inventor and invention dates of each claim that was not commonly owned at the time a later invention was made in order for the examiner to consider the applicability of 35 U.S.C. 103(c) and potential 35 U.S.C. 102(e), (f) or (g) prior art under 35 U.S.C. 103(a).

10. Claims 61 and 66 are rejected under 35 U.S.C. 103(a) as being unpatentable over Caspari et al (US Pat. No.: 4,923,461).

In regard to Claims 61 and 66, Caspari et al discloses all the limitations of the claims but fails to disclose the first elongated member comprises an eyelet. However, it is old and well known in the art that an elongated tissue-puncturing member / a suturing passing member / a needle has an eyelet to carry or manipulate a suture strand. Therefore, it would have been obvious to modify the needle of Caspari et al to include an eyelet for carrying and manipulating the suture.

11. Claims 53, 56 and 57 are rejected under 35 U.S.C. 103(a) as being unpatentable over Caspari et al (US Pat. No.: 4,923,461) in view of Middleman et al (US Pat. No.:5,820,628).

In regard to Claims 53, 56, and 57, Caspari et al discloses a surgical device (suturing instrument, 20, Fig. 1) for passing suture through soft tissue, comprising: a grasping mechanism comprising a first jaw (44, Fig. 2) and second jaw (34, Fig. 2) connected to said body and encompassing at least a portion of the elongate member, one of the jaws moveable relative to the other and having an opening (65, Fig. 1) at the middle region of the jaw to receive at least one strand of suture for allowing the elongate member therethrough and to the soft tissue, and wherein said grasping mechanism clamps soft tissue while the suture passing member slides within the lumen and passes suture through the soft tissue outside of the grasping mechanism through an opening in one of said jaws of said grasping mechanism (Figs. 11-12 and Col. 4, lines 59-68); a second elongate member (elongate tube, 36, Fig. 1) having a lumen to accommodate a

first elongate member (combination of tubing 104 and needle 56, Figs 5 and 6 and Col. 4, lines 19-22).

Caspari et al does not appear to disclose that the first elongate member has a first resting configuration, second stressed configuration and an eyelet.

However, Middleman et al explicitly discloses a needle/elongate member has a first resting configuration (curved configuration, Fig. 15C) and is compressible into a second, stressed configuration (relatively straight configuration, 15A), and returns towards said first resting configuration as a compressive external force is reduced (Figs. 15A-C and Col. 26, lines 26-40).

It is old and well-known in the art to include a needle having the abilities of obtaining a compressed configuration in a delivery tube and return to its resting configuration outside the delivery tube. The motivation/suggestion for do so would have been easier to deliver such needle within a straight elongate body to a target site in the human body. In addition, it is old and well known in the art that a needle has an eyelet to carry or manipulate a suture strand. Therefore, it would have been obvious to modify the needle of Caspari et al to include an eyelet for carrying and manipulating the suture.

### ***Response to Arguments***

12. Applicant's arguments with respect to claims 50, 53, 56, 57, 59-62, and 65- 67 have been considered but are moot in view of the new ground(s) of rejection.

13. Prosecution on the merits of this application is re-opened, all previous indicated allowable claims are withdrawn.

***Conclusion***

Any inquiry concerning this communication or earlier communications from the examiner should be directed to JING OU whose telephone number is (571)270-5036. The examiner can normally be reached on M-F 7:30am - 5:00pm, Alternative Friday.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Uyen (Jackie) T Ho can be reached on (571)272-4696. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

JO

/Darwin P. Erezio/  
Primary Examiner, Art Unit 3773



Application/Control Number: 10/084,283  
Art Unit: 3773

Page 8